

REPORT ON BIO-MEDICAL WASTE MANAGEMENT FOR THE MONTH OF MARCH, 2022

Sl. No.	Description	Remarks
1	Contact Address and details of the Healthcare Facility	LGB Regional Institute of Mental Health, Tezpur, Sonitpur, Assam, Pin-784001, Tele No. 03712-232652, fax No.- 03712-233623
2	No. of Beds Details of :	336 No.
3	a) Authorisation under BMW Rules, 2016 b) Consent under Water (Prevention and Control of Pollution) Act, 1974 and Air(Prevention and Control of Pollution) Act, 1981	Applied for, Application no: 444804 dated, 12/03/2021 Previous authorization No: WB/OTWA/BMW-448/18-19/01/170 Applied for, Application no: 444804, dated, 12/03/2021 Previous authorization No: WB/TEZT-1745/18-19/01/1489
4	Quantity of bio-medical waste generation (in Kg/day)	7.59 Kg
5	Mode of disposal of bio-medical waste (through CBWTF or through captive treatment facility)	Through captive treatment facility
6	Name and address of the CBWTF through which waste is disposed off (as applicable) In case, HCF is having captive treatment facility.	N/A
7	a) Bio-medical waste treated (in Kg/day)	7.59 Kg
	b) Details of treatment equipment	i) Incinerator ii) Autoclave iii) Shredder iv) Needle Destroyer v) ETP (Effluent Treatment Plant)
	c) Total nos. And capacity of each treatment equipment (in Kg/day)	i) Incinerator- 1 No. Capacity: 10-12 Kg/hr ii) Autoclave - 1 No. Capacity : 78 litres iii) Shredder - 1 No. Capacity : 50 -100 kg iv) Needle Destroyer - 12 Nos. Capacity: — v) ETP - 1 No. Capacity: 4000 litre/day

  
 IIC  
 Central Laboratory  
 LGB, RIMH  
 Tezpur

9	No. of trainings conducted on Bio-medical Waste management in the current year	28																		
8	Monthly records of biomedical waste generation(category wise)	<p>i) Yellow category : 71,322 kg/month  ii) Red category : 120,424 kg/month  iii) White category : 7,223 kg/month  iv) Blue category : 35,325 kg/month</p>																		
	<p>d) Operating parameters of the treatment equipment as per BMW/M Rules, 2016</p>	<p><b>1) Incinerator</b></p> <table border="1" data-bbox="1113 1366 1317 2060"> <thead> <tr> <th>Sl. No.</th> <th>Parameter</th> <th>Result</th> </tr> </thead> <tbody> <tr> <td>i.</td> <td>SO<sub>2</sub> (ppm)</td> <td>17.5</td> </tr> <tr> <td>ii.</td> <td>NO<sub>x</sub> (ppm)</td> <td>37.5</td> </tr> <tr> <td>iii.</td> <td>HCl (mg/Nm<sup>3</sup>)</td> <td>29</td> </tr> <tr> <td>iv.</td> <td>CO (ppm)</td> <td>286</td> </tr> <tr> <td>v.</td> <td>CO<sub>2</sub> %</td> <td>8.2</td> </tr> </tbody> </table> <p><b>Temperature</b>  Primary Chamber 800° ± 50° C,  Secondary Chamber 1000° ± 50° C  <b>Height</b>- 30 meters above from the ground  <b>Capacity</b>- 10-12 Kg/hr</p> <p><b>2) Autoclave: (Gravite flow)</b></p> <p>i) Monitors steam processes with Geobacillus stearothermophilus spores.  ii) Steam Indicator Tape</p>	Sl. No.	Parameter	Result	i.	SO <sub>2</sub> (ppm)	17.5	ii.	NO <sub>x</sub> (ppm)	37.5	iii.	HCl (mg/Nm <sup>3</sup> )	29	iv.	CO (ppm)	286	v.	CO <sub>2</sub> %	8.2
Sl. No.	Parameter	Result																		
i.	SO <sub>2</sub> (ppm)	17.5																		
ii.	NO <sub>x</sub> (ppm)	37.5																		
iii.	HCl (mg/Nm <sup>3</sup> )	29																		
iv.	CO (ppm)	286																		
v.	CO <sub>2</sub> %	8.2																		

  
01/04/22  
I/C Central Laboratory, IIC  
LGBRIMIL, Tezpur  
Central Laboratory  
LGBRIMH  
Tezpur